



測定値 (mg/L)	1回目	116.8	121.5	123.1	126.4	129.8	136.0	132.8	133.8	140.9	147.0	149.3	140.7	153.9	157.4	157.2	169.8	160.9	156.7	162.2	160.5	158.8	164.1	163.5	166.0	170.6	167.2	174.3	172.2	169.4	180.4	177.0	179.2	229.1	
	2回目	115.3	120.9	124.8	131.3	129.8	132.0	135.2	127.7	141.3	146.6	143.2	151.9	148.0	154.6	156.0	148.1	153.5	160.4	153.1	160.4	164.8	166.6	163.8	167.0	164.3	164.5	165.0	169.1	172.2	172.5	180.0	182.4	230.8	
	3回目	118.7	113.9	122.4	127.7	135.2	132.0	134.4	142.8	137.3	140.4	146.2	148.0	149.4	154.6	154.8	153.7	157.2	157.1	160.8	159.6	162.8	161.1	165.8	167.3	165.6	168.9	171.3	170.1	173.0	166.0	177.0	177.0	181.2	234.1
平均値(mg/L)①	116.9	118.8	123.4	128.5	131.6	133.3	134.1	134.8	139.8	144.7	146.2	146.9	150.4	155.5	156.0	157.2	157.2	158.1	158.7	160.2	162.1	163.9	164.4	166.8	166.8	166.9	170.2	170.5	171.5	173.0	178.0	181.2	231.3		
標準偏差②	1.70	4.23	1.23	2.54	3.12	2.31	1.22	7.60	2.20	3.70	3.05	5.69	2.31	3.08	1.62	1.20	11.27	3.70	2.03	4.90	0.49	3.06	2.75	1.25	0.68	3.33	2.22	4.75	1.58	1.89	7.21	1.73	1.77	2.54	
CV(室内)(%) (②/①)	1.46	3.56	1.00	1.98	2.37	1.73	0.91	5.64	1.58	2.56	2.09	3.87	2.05	1.04	0.77	7.17	2.35	1.28	3.09	0.31	1.88	1.68	0.76	2.35	0.41	1.99	1.33	2.79	0.93	1.10	4.17	0.97	0.98	1.10	
参加者平均との差	-35.80	-33.97	-29.30	-24.27	-21.13	-19.40	-18.60	-17.97	-12.90	-8.07	-6.50	-5.87	-2.30	2.80	3.27	4.47	4.47	5.33	5.97	7.43	9.40	11.20	11.63	14.03	14.10	14.13	17.47	17.73	18.80	20.23	25.27	28.50	78.60		
Zスコア	-2.03	-1.93	-1.66	-1.38	-1.20	-1.10	-1.06	-1.02	-0.73	-0.46	-0.37	-0.33	-0.13	0.16	0.18	0.25	0.25	0.30	0.34	0.42	0.53	0.63	0.66	0.80	0.80	0.80	0.99	1.00	1.07	1.15	1.43	1.62	4.46		
Grubbs検定による棄却																																			○

参加者平均	152.7
標準偏差	17.64
CV(室内)%	2.06

Zスコア	事業所数	33
$ Z \leq 2$		31
$2 < Z < 3$		1
$3 \leq Z $		1